

Deutsche Akkreditierungsstelle GmbH

Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV

Signatory to the Multilateral Agreements of EA, ILAC and IAF for Mutual Recognition

Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

Gigahertz-Optik GmbH An der Kälberweide 12 82299 Türkenfeld

is competent under the terms of DIN EN ISO/IEC 17025:2005 to carry out calibrations in the following fields:

Optical Quantities

Radiometry

The accreditation certificate shall only apply in connection with the notice of accreditation of 2010-10-11 with the accreditation number D-K-15047-01 and is valid until 2013-02-28. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 2 pages.

Registration number of the certificate: D-K-15047-01-00

Braunschweig, 2010-10-11

by order of Dr. Michael Wolf Head of Division 5

This document is a translation. The definitive version is the original German accreditation certificate.

Deutsche Akkreditierungsstelle GmbH

Office Berlin Spittelmarkt 10 10117 Berlin Office Frankfurt am Main Gartenstraße 6 60594 Frankfurt am Main Office Braunschweig Bundesallee 100 38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkkS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkkS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkkS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: www.european-accreditation.org

ILAC: www.ilac.org IAF: www.iaf.nu



Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D- K-15047-01-00 according to DIN EN ISO/IEC 17025:2005

Period of validity: 2010-10-11 to 2013-02-28

Holder of certificate:

Gigahertz-Optik GmbH An der Kälberweide 12 82299 Türkenfeld

Head:

Dipl.-Ing. (FH) Anton Gugg-Helminger

Deputy:

Dipl.-Ing. (FH) Stephan Fenk

Accredited since:

1993-06-22

Calibrations in the fields:

Optical Quantities

Radiometry



Annex to the accreditation certificate D-K-15047-01-00

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability 1)	Remarks
Spectral responsivity of photodiodes	Wavelength 248.3 nm; 265.3 nm; 280.4 nm; 289.4 nm; 302.2 nm; 313 nm; 334.2 nm; 366 nm	$1 \text{ nW/cm}^2 \leq E \leq 10 \mu\text{W/cm}^2$ $18 \text{ °C} \leq t \leq 28 \text{ °C}$ $1 \text{ nm} \leq \Delta \lambda \leq 11 \text{ nm}$	3 %	Measurement of photocurrent by means of a trans-impedance amplifier with the terminal voltage of the photodiode $U \le 50 \mu V$. $E = irradiance$ at the area of the detector
Spectral responsivity as a function of the wavelength	380 nm to 900 nm steps of 20 nm	100 nW/cm ² $\leq E \leq$ 100 μW/cm ² 18 °C $\leq t \leq$ 28 °C 1 nm $\leq \Delta \lambda \leq$ 11 nm	2 %	
	900 nm to 1040 nm steps of 20 nm		2.5 %	t = ambient tempera- ture of the calibration item
	1060 nm; 1080 nm; 1100 nm		3 %	$\Delta \lambda$ = full width at half maximum
	1120 nm; 1140 nm; 1160 nm		4 %	
Spectral irradiance		Wavelength λ :		P = electrical power Calibration with source standards traceable to
of lamps (thermal radiators)	500 W ≤ <i>P</i> ≤ 2000 W	250 nm	10 %	
		260 nm	7 %	
		270 nm $\leq \lambda <$ 400 nm	4 %	PTB standards at wave
		400 nm $\leq \lambda < 800$ nm	3 %	lengths equal to those
		800 nm $\leq \lambda < 2000$ nm	4.5 %	used for the calibration
		2000 nm $\leq \lambda \leq$ 2500 nm	7 %	of these standards.
		250 nm	12 %	Calibration with
		260 nm	8 %	internally calibrated working standards at wavelength equal to those used for the
		270 nm ≤ λ < 400 nm	4.5 %	
		400 nm ≤ λ < 800 nm	3.5 %	
		800 nm $\leq \lambda < 2000$ nm	5.5 %	
		2000 nm ≤ λ ≤ 2500 nm	8.5 %	calibration of these standards.

¹⁾ The table shows the best measurement capabilities expressed as expanded uncertainties of measurement according to DAkkS-DKD-3 (EA-4/02). The expanded uncertainties of measurement correspond to a coverage probability of 95% and have a coverage factor of k = 2 unless stated otherwise. Uncertainties without unit are relative uncertainties unless stated otherwise.